


RaySafe Xi – Toshiba holder






Mount the holder straight or diagonally depending on the dimensions of the image receptor.



12x12, 12x16 and Kalare



8x8



Note: Make sure the RaySafe Xi detector is positioned perpendicular to the X-ray tube axis.

