



# DECLARATION OF CONFORMITY

Product Identification:

First declared: 2019

**Fluke Biomedical RaySafe 452 - Radiation Survey Meter Family**

Including Models: 452, 452 Ambient, 452 Air Kerma

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO 9001:2015 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

**Directive 2015/863/EU, UK Regulation 2016: RoHS**

**Directive 2014/30/EU, UK Regulation 2016: Electromagnetic Compatibility (EMC)**

**Directive 2014/35/EU, UK Regulation 2016: Low Voltage (LVD)**

Standards Used:

EN/BS 63000:2016 Assessment of electrical and electronic products for restriction of hazardous substances

EN/BS 61326-1:2013 Electrical equipment for measurement, control and laboratory-EMC; General

EN/BS 61010-1:2010 Safety requirements for measurement, control and laboratory equipment; General

EN/BS 60529: 1991 +A2:2013 Degrees of protection provided by enclosures

-----

Special Conditions:

None

EU Importer/Distributor

Fluke Europe B.V.  
PO Box 1186  
5602 BD Eindhoven  
The Netherlands

UK Importer/Distributor

Fluke Precision Measurement Ltd.  
52 Hurricane Way  
Norwich, NR6 6JB  
United Kingdom

Manufacturer

Fluke Corporation  
6920 Seaway Boulevard M/S 143F  
Everett, WA 98203; USA

Henry Kuster  
EMEA Logistics and Compliance Manager  
Fluke Europe B.V.

Thomas M. Smith P.E.  
Product Compliance Manager  
Fluke Corporation