



EU DECLARATION OF CONFORMITY

Product Identification:

Fluke Biomedical RaySafe 452 - Radiation Survey Meter Family
Including Models: 452, 452 Ambient, 452 Air Kerma

First declared: 2019

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO 9001:2015 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

Directive 2011/65/EU, RoHS
Directive 2014/30/EU, Electromagnetic Compatibility (EMC)
Directive 2014/35/EU, Low Voltage (LVD)

Standards Used:

EN 50581: 2012 Restriction of the use of hazardous substances
EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General
EN 61010-1: 2010 Safety requirements for measurement, control and laboratory; General
EN 60529: 1991 +A2:2013 Degrees of protection provided by enclosures


Special Conditions:

None

Manufacturer

Fluke Biomedical
6920 Seaway Boulevard M/S 143F
Everett, WA 98203; USA

CE-111 Rev.001


Thomas M. Smith P.E.

Product Compliance Manager, Fluke Corporation

CE-111 Rev.001 - This declaration supersedes all previous declarations for this product.

29 April 2019